

Annex 4 – Calibration/ Conformance Certificates for Sound Level Meters and Calibrator



CERTIFICATE OF CALIBRATION



0653

CAL 001

Date of Issue: 02 January 2018

Certificate Number: UCRT18/1003

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

| |
|--------------------|
| Page 1 of 2 Pages |
| Approved Signatory |
| |
| K. Mistry |

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Customer TNEI Services Ltd
 Milburn House
 Dean Street
 Newcastle Upon Tyne
 NE1 1LE

Order No. 5001

Test Procedure Procedure TP 1 Calibration of Sound Calibrators

Description Acoustic Calibrator

| Identification | Manufacturer | Instrument | Model | Serial No. |
|----------------|--------------|------------|-------|------------|
| | Rion | Calibrator | NC-74 | 34762316 |

The calibrator has been tested as specified in Annex B of IEC 60942:2003. As public evidence was available from a testing organisation (PTB) responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound calibrator fully conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, the sound calibrator tested is considered to conform to all the class 1 requirements of IEC 60942:2003.

ANV Job No. UKAS17/12696

Date Received 21 December 2017

Date Calibrated 02 January 2018

Previous Certificate

| | |
|-----------------|-----------------|
| Dated | 24 January 2017 |
| Certificate No. | UCRT17/1033 |
| Laboratory | 7623 |

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate Number

UCRT18/1003

UKAS Accredited Calibration Laboratory No. 0653

Page 2 of 2 Pages

Measurements

The sound pressure level generated by the calibrator in its WS2 configuration was measured five times by the Insert Voltage Method using a microphone as detailed below. The mean of the results obtained is shown below. It is corrected to the standard atmospheric pressure of 101.3 kPa (1013 mBar) using original manufacturers information.

| Test Microphone | Manufacturer | Type |
|-----------------|--------------|------|
| | Brüel & Kjær | 4134 |

Results

The level of the calibrator output under the conditions outlined above was

94.00 ± 0.10 dB rel 20 µPa

Functional Tests and Observations

| | | | |
|---|------------|---|------------------|
| The frequency of the sound produced was | 1002.49 Hz | ± | 0.13 Hz |
| The total distortion was | 1.51 % | ± | 6.6 % of Reading |

During the measurements environmental conditions were

| | | | |
|---------------------|------|----|----------|
| Temperature | 22 | to | 22 °C |
| Relative Humidity | 38 | to | 48 % |
| Barometric Pressure | 99.2 | to | 99.4 kPa |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The uncertainties refer to the measured values only with no account being taken of the ability of the instrument to maintain its calibration.

A small correction factor may need to be applied to the sound pressure level quoted above if the device is used to calibrate a sound level meter which is fitted with a free-field response microphone. See manufacturers handbook for details.

..... END

Note:

| | |
|---|--------|
| Calibrator adjusted prior to calibration? | NO |
| Initial Level | N/A dB |
| Initial Frequency | N/A Hz |

Additional Comments

None

Calibrated by: B. Bogdan

R 1



CERTIFICATE OF CALIBRATION

Date of Issue: 05 September 2017

Certificate Number: TCRT17/1572

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Page 1 of 2 Pages
Approved Signatory

K. Mistry

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Customer TNEI Services Ltd
Milburn House
Dean Street
Newcastle Upon Tyne
NE1 1LE

Order No. 5001

Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator

Identification

| Manufacturer | Instrument | Type | Serial No. / Version |
|--------------|---------------------------------------|-------|----------------------|
| Rion | Sound Level Meter | NL-52 | 00643025 |
| Rion | Firmware | | 1.8 |
| Rion | Pre Amplifier | NH-25 | 43053 |
| Rion | Microphone | UC-59 | 06805 |
| Rion | Calibrator | NC-74 | 34536109 |
| | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic tests.

Type Approved to IEC 61672-1:2002 YES Approval Number 21.21 / 13.02

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 04 September 2017

ANV Job No. TRAC17/09334

Date Calibrated 05 September 2017

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. As public evidence was available, from an independent testing organisation responsible for approving the results of pattern evaluation tests performed in accordance with IEC 61672-2:2003, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002, the sound level meter submitted for testing conforms to the class 1 requirements of IEC 61672-1:2002.

Previous Certificate

Dated

Certificate No.

Laboratory

15 September 2016

TCRT16/1248

ANV Measurement Systems

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT17/1578

Page 2 of 2 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | |
|--|-------------------|---|
| SLM instruction manual title | Sound Level Meter | NL-42 / NL-52 |
| SLM instruction manual ref / issue | | 11-03 |
| SLM instruction manual source | Manufacturer | |
| Internet download date if applicable | | N/A |
| Case corrections available | Yes | |
| Uncertainties of case corrections | Yes | |
| Source of case data | Manufacturer | |
| Wind screen corrections available | Yes | |
| Uncertainties of wind screen corrections | Yes | |
| Source of wind screen data | Manufacturer | |
| Mic pressure to free field corrections | Yes | |
| Uncertainties of Mic to F.F. corrections | Yes | |
| Source of Mic to F.F. corrections | Manufacturer | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | | Yes |
| Specified or equivalent Calibrator | Specified | |
| Customer or Lab Calibrator | Lab Calibrator | |
| Calibrator adaptor type if applicable | NC-74-002 | |
| Calibrator cal. date | 21 August 2017 | |
| Calibrator cert. number | UCRT17/1705 | |
| Calibrator cal cert issued by | 0653 | |
| Calibrator SPL @ STP | 94.04 | dB Calibration reference sound pressure level |
| Calibrator frequency | 1001.94 | Hz Calibration check frequency |
| Reference level range | 25 - 130 | dB |

Accessories used or corrected for during calibration - None
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 22.24 | 22.37 | ± 0.20 °C |
| Humidity | 43.6 | 43.1 | ± 3.00 %RH |
| Ambient Pressure | 100.84 | 100.84 | ± 0.03 kPa |

| | | | |
|--|------|----|----------------------------------|
| Response to associated Calibrator at the environmental conditions above. | | | |
| Initial indicated level | 93.9 | dB | Adjusted indicated level 94.0 dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB |

| | | | | | | |
|--|---|-------|-------------|-------|------|-------|
| Self Generated Noise | This test is currently not performed by this Lab. | | | | | |
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting | | | |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | | | | |
| Microphone replaced with electrical input device - | UR = Under Range indicated | | | | | |
| Weighting | A | | C | | Z | |
| | 12.3 | dB UR | 17.1 | dB UR | 22.9 | dB UR |
| Uncertainty of the electrical self generated noise ± | | | 0.12 | | dB | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by ISO.

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

..... END

Calibrated by: A Patel

Additional Comments

None



SL11

CERTIFICATE OF CALIBRATION

Date of Issue: 01 December 2017

Certificate Number: TCRT17/1794

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 3 Pages

Approved Signatory

K. Mistry

Customer TNEI Services Ltd
 Milburn House
 Dean Street
 Newcastle Upon Tyne
 NE1 1LE

Order No. 5001
 Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator
 Identification

| Manufacturer | Instrument | Type | Serial No. / Version |
|--------------|---------------------------------------|--------|----------------------|
| Rion | Sound Level Meter | NL-31 | 01273082 |
| Rion | Firmware | | 1.400 |
| Rion | Pre Amplifier | NH-21 | 26001 |
| Rion | Microphone | UC-53A | 313385 |
| Rion | Calibrator | NC-74 | 34536109 |
| | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic test.

Type Approved to IEC 61672-1:2002 No Approval Number

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 28 November 2017 ANV Job No. TRAC17/11472

Date Calibrated 01 December 2017

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002 and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|-----------------|-----------------|-------------------------|
| | 16 October 2015 | TCRT15/1274 | ANV Measurement Systems |

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT17/1794

Page 2 of 3 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|--------------------------------|-----------------------|--|
| SLM instruction manual title | NL-21 NL-31 Instruction Manual | | |
| SLM instruction manual ref / issue | 32006 09-04 | | |
| SLM instruction manual source | Manufacturer | | |
| Internet download date if applicable | N/A | | |
| Case corrections available | Yes | | |
| Uncertainties of case corrections | No | See comment on page 3 | |
| Source of case data | Manufacturer | | |
| Wind screen corrections available | Yes | | |
| Uncertainties of wind screen corrections | No | See comment on page 3 | |
| Source of wind screen data | Manufacturer | | |
| Mic pressure to free field corrections | Yes | | |
| Uncertainties of Mic to F.F. corrections | No | See comment on page 3 | |
| Source of Mic to F.F. corrections | Manufacturer | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Lab Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 13 November 2017 | | |
| Calibrator cert. number | UCRT17/2023 | | |
| Calibrator cal cert issued by Lab. | 0653 | | |
| Calibrator SPL @ STP | 94.01 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1001.95 | Hz | Calibration check frequency |
| Reference level range | 30 - 120 | dB | |

Accessories used or corrected for during calibration - None
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 22.20 | 22.70 | ± 0.30 °C |
| Humidity | 41.2 | 41.5 | ± 3.00 %RH |
| Ambient Pressure | 101.19 | 101.23 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|--|------|----|--------------------------|------|----|
| Initial indicated level | 94.1 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|--|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

Microphone replaced with electrical input device - UR = Under Range indicated

| Weighting | A | | | C | | | Z | | | |
|--|-----|----|--|------|----|--|------|----|--|--|
| | dB | UR | | dB | UR | | dB | UR | | |
| | 9.6 | | | 16.4 | | | 21.9 | | | |
| Uncertainty of the electrical self generated noise ± | | | | | | | 0.12 | dB | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by the International Organisation for Standards (ISO).

Comments

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT17/1794

Page 3 of 3 Pages

If any of the "Uncertainties of" are set to NO above, then the following applies.

No information on the uncertainty of measurement, required by 11.7 of IEC 61672-3:2006, of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, or the manufacturer of the multi-frequency sound calibrator, or the manufacturer of the electrostatic actuator was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of the measurement of the adjustment data has therefore been assumed to be numerically zero for the purpose of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of IEC 61672-1:2002.

Calibrated by: B. Bogdan

R 1

..... END

Additional Comments

None



CERTIFICATE OF CALIBRATION

Date of Issue: 01 December 2017

Certificate Number: TCRT17/1793

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 3 Pages

Approved Signatory

K. Mistry

Customer TNEI Services Ltd
Milburn House
Dean Street
Newcastle Upon Tyne
NE1 1LE

| | | | | |
|----------------|--|---------------------------------------|-------------|-----------------------------|
| Order No. | 5001 | | | |
| Description | Sound Level Meter / Pre-amp / Microphone / Associated Calibrator | | | |
| Identification | <i>Manufacturer</i> | <i>Instrument</i> | <i>Type</i> | <i>Serial No. / Version</i> |
| | Rion | Sound Level Meter | NL-31 | 00593593 |
| | Rion | Firmware | | 1.400 |
| | Rion | Pre Amplifier | NH-21 | 30355 |
| | Rion | Microphone | UC-53A | 316118 |
| | Rion | Calibrator | NC-74 | 34536109 |
| | | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic test.

Type Approved to IEC 61672-1:2002 No Approval Number

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 28 November 2017

ANV Job No. TRAC17/11472

Date Calibrated 01 December 2017

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002 and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

| | | | |
|----------------------|------------------|------------------------|-------------------------|
| Previous Certificate | <i>Dated</i> | <i>Certificate No.</i> | <i>Laboratory</i> |
| | 16 February 2015 | TCRT15/1055 | ANV Measurement Systems |

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT17/1793

Page 2 of 3 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|--------------------------------|-----------------------|--|
| SLM instruction manual title | NL-21 NL-31 Instruction Manual | | |
| SLM instruction manual ref / issue | 32006 09-04 | | |
| SLM instruction manual source | Manufacturer | | |
| Internet download date if applicable | N/A | | |
| Case corrections available | Yes | | |
| Uncertainties of case corrections | No | See comment on page 3 | |
| Source of case data | Manufacturer | | |
| Wind screen corrections available | Yes | | |
| Uncertainties of wind screen corrections | No | See comment on page 3 | |
| Source of wind screen data | Manufacturer | | |
| Mic pressure to free field corrections | Yes | | |
| Uncertainties of Mic to F.F. corrections | No | See comment on page 3 | |
| Source of Mic to F.F. corrections | Manufacturer | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Lab Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 13 November 2017 | | |
| Calibrator cert. number | UCRT17/2023 | | |
| Calibrator cal cert issued by Lab. | 0653 | | |
| Calibrator SPL @ STP | 94.01 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1001.95 | Hz | Calibration check frequency |
| Reference level range | 30 - 120 | dB | |

Accessories used or corrected for during calibration - None

Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 21.70 | 22.30 | ± 0.30 °C |
| Humidity | 41.7 | 40.4 | ± 3.00 %RH |
| Ambient Pressure | 101.19 | 101.19 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|--|------|----|--------------------------|------|----|
| Initial indicated level | 94.3 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|--|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

Microphone replaced with electrical input device - UR = Under Range indicated

| Weighting | A | | C | | Z | |
|--|------|----|------|----|------|----|
| | dB | UR | dB | UR | dB | UR |
| | 10.9 | UR | 17.6 | UR | 22.5 | UR |
| Uncertainty of the electrical self generated noise ± | 0.12 | | dB | | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by the International Organisation for Standards (ISO).

Comments

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT17/1793

Page 3 of 3 Pages

If any of the "Uncertainties of" are set to NO above, then the following applies.

No information on the uncertainty of measurement, required by 11.7 of IEC 61672-3:2006, of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, or the manufacturer of the multi-frequency sound calibrator, or the manufacturer of the electrostatic actuator was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of the measurement of the adjustment data has therefore been assumed to be numerically zero for the purpose of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of IEC 61672-1:2002.

Calibrated by: B. Bogdan

R 1

..... END

Additional Comments

None



CERTIFICATE OF CALIBRATION

Date of Issue: 14 March 2018

Certificate Number: TCRT18/1213

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 3 Pages

Approved Signatory

K. Mistry

Customer TNEI Services Ltd
Milburn House
Dean Street
Newcastle Upon Tyne
NE1 1LE

Order No. 5001
Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator
Identification

| Manufacturer | Instrument | Type | Serial No. / Version |
|--------------|---------------------------------------|--------|----------------------|
| Rion | Sound Level Meter | NL-31 | 00593595 |
| Rion | Firmware | | 1.400 |
| Rion | Pre Amplifier | NH-21 | 30357 |
| Rion | Microphone | UC-53A | 316120 |
| Rion | Calibrator | NC-74 | 34973250 |
| | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic test.

Type Approved to IEC 61672-1:2002 No Approval Number

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 13 March 2018

ANV Job No. TRAC18/03118

Date Calibrated 14 March 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002 and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|---------------|-----------------|-------------------------|
| | 10 March 2017 | TCRT17/1076 | ANV Measurement Systems |

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1213

Page 2 of 3 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|--------------------------------|-----------------------|--|
| SLM instruction manual title | NL-21 NL-31 Instruction Manual | | |
| SLM instruction manual ref / issue | 32006 09-04 | | |
| SLM instruction manual source | Manufacturer | | |
| Internet download date if applicable | N/A | | |
| Case corrections available | Yes | | |
| Uncertainties of case corrections | No | See comment on page 3 | |
| Source of case data | Manufacturer | | |
| Wind screen corrections available | Yes | | |
| Uncertainties of wind screen corrections | No | See comment on page 3 | |
| Source of wind screen data | Manufacturer | | |
| Mic pressure to free field corrections | Yes | | |
| Uncertainties of Mic to F.F. corrections | No | See comment on page 3 | |
| Source of Mic to F.F. corrections | Manufacturer | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Customers Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 14 March 2018 | | |
| Calibrator cert. number | UCRT18/1293 | | |
| Calibrator cal cert issued by Lab. | 0653 | | |
| Calibrator SPL @ STP | 93.99 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1002.89 | Hz | Calibration check frequency |
| Reference level range | 30 - 120 | dB | |

Accessories used or corrected for during calibration - Wind Shield WS-10
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|-------|-------|------------|
| Temperature | 22.87 | 22.70 | ± 0.30 °C |
| Humidity | 44.0 | 45.5 | ± 3.00 %RH |
| Ambient Pressure | 99.21 | 99.11 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|--|------|----|--------------------------|------|----|
| Initial indicated level | 94.3 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|--|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

Microphone replaced with electrical input device - UR = Under Range indicated

| Weighting | A | | | C | | | Z | | | |
|--|------|----|--|------|----|--|------|----|--|--|
| | dB | UR | | dB | UR | | dB | UR | | |
| | 11.1 | | | 17.1 | | | 23.4 | | | |
| Uncertainty of the electrical self generated noise ± | | | | | | | 0.12 | dB | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by the International Organisation for Standards (ISO).

Comments

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1213

Page 3 of 3 Pages

If any of the "Uncertainties of" are set to NO above, then the following applies.

No information on the uncertainty of measurement, required by 11.7 of IEC 61672-3:2006, of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, or the manufacturer of the multi-frequency sound calibrator, or the manufacturer of the electrostatic actuator was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of the measurement of the adjustment data has therefore been assumed to be numerically zero for the purpose of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of IEC 61672-1:2002.

Calibrated by: B. Bogdan

R 2

..... END

Additional Comments

None

SLM001



CERTIFICATE OF CALIBRATION

Date of Issue: 16 February 2018

Certificate Number: TCRT18/1163

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 3 Pages

Approved Signatory

K. Mistry

Customer TNEI Services Ltd
Milburn House
Dean Street
Newcastle Upon Tyne
NE1 1LE

| | | | | |
|----------------|--|---------------------------------------|-------------|-----------------------------|
| Order No. | 5001 | | | |
| Description | Sound Level Meter / Pre-amp / Microphone / Associated Calibrator | | | |
| Identification | <i>Manufacturer</i> | <i>Instrument</i> | <i>Type</i> | <i>Serial No. / Version</i> |
| | Rion | Sound Level Meter | NL-32 | 00661767 |
| | Rion | Firmware | | 1.0009 |
| | Rion | Pre Amplifier | NH-21 | 19771 |
| | Rion | Microphone | UC-53A | 310458 |
| | Rion | Calibrator | NC-74 | 34536109 |
| | | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic test.

Type Approved to IEC 61672-1:2002 No Approval Number

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 15 February 2018

ANV Job No. TRAC18/02083

Date Calibrated 16 February 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002 and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

| | | | |
|----------------------|-----------------|------------------------|-------------------------|
| Previous Certificate | <i>Dated</i> | <i>Certificate No.</i> | <i>Laboratory</i> |
| | 25 January 2017 | TCRT17/1024 | ANV Measurement Systems |

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1163

Page 2 of 3 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|--------------------------------|-----------------------|--|
| SLM instruction manual title | NL-22 NL-32 Instruction Manual | | |
| SLM instruction manual ref / issue | 33625 09-06 | | |
| SLM instruction manual source | Manufacturer | | |
| Internet download date if applicable | N/A | | |
| Case corrections available | Yes | | |
| Uncertainties of case corrections | No | See comment on page 3 | |
| Source of case data | Manufacturer | | |
| Wind screen corrections available | Yes | | |
| Uncertainties of wind screen corrections | No | See comment on page 3 | |
| Source of wind screen data | Manufacturer | | |
| Mic pressure to free field corrections | Yes | | |
| Uncertainties of Mic to F.F. corrections | No | See comment on page 3 | |
| Source of Mic to F.F. corrections | Manufacturer | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Lab Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 07 February 2018 | | |
| Calibrator cert. number | UCRT18/1145 | | |
| Calibrator cal cert issued by Lab. | 0653 | | |
| Calibrator SPL @ STP | 94.03 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1001.88 | Hz | Calibration check frequency |
| Reference level range | 30 - 120 | dB | |

Accessories used or corrected for during calibration - None

Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 22.18 | 22.45 | ± 0.30 °C |
| Humidity | 40.0 | 40.8 | ± 3.00 %RH |
| Ambient Pressure | 101.47 | 101.42 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|--|------|----|--------------------------|------|----|
| Initial indicated level | 94.2 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | | 0.10 | dB |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|--|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

| | | | | | | |
|--|----------------------------|----|------|------|----|----|
| Microphone replaced with electrical input device - | UR = Under Range indicated | | | | | |
| Weighting | A | | C | | Z | |
| | 10.9 | dB | UR | 17.8 | dB | UR |
| Uncertainty of the electrical self generated noise ± | | | 0.12 | | | dB |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by the International Organisation for Standards (ISO).

Comments

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1163

Page 3 of 3 Pages

If any of the "Uncertainties of" are set to NO above, then the following applies.

No information on the uncertainty of measurement, required by 11.7 of IEC 61672-3:2006, of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, or the manufacturer of the multi-frequency sound calibrator, or the manufacturer of the electrostatic actuator was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of the measurement of the adjustment data has therefore been assumed to be numerically zero for the purpose of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of IEC 61672-1:2002.

Calibrated by: B. Bogdan

R 2

..... END

Additional Comments

None

SLM9



CERTIFICATE OF CALIBRATION

Date of Issue: 16 February 2018

Certificate Number: TCRT18/1165

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 3 Pages

Approved Signatory

K. Mistry

Customer TNEI Services Ltd
Milburn House
Dean Street
Newcastle Upon Tyne
NE1 1LE

| | | | | |
|-----------------------|--|---------------------------------------|-------------|-----------------------------|
| Order No. | 5001 | | | |
| Description | Sound Level Meter / Pre-amp / Microphone / Associated Calibrator | | | |
| Identification | <i>Manufacturer</i> | <i>Instrument</i> | <i>Type</i> | <i>Serial No. / Version</i> |
| | Rion | Sound Level Meter | NL-32 | 00972337 |
| | Rion | Firmware | | 1.0009 |
| | Rion | Pre Amplifier | NH-21 | 25122 |
| | Rion | Microphone | UC-53A | 313228 |
| | Rion | Calibrator | NC-74 | 34536109 |
| | | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1

Test Procedure TP 2.SLM 61672-3 TPS-49

Procedures from IEC 61672-3:2006 were used to perform the periodic test.

Type Approved to IEC 61672-1:2002 No **Approval Number**

If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003

Date Received 15 February 2018

ANV Job No. TRAC18/02083

Date Calibrated 16 February 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002 and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

| | | | |
|-----------------------------|-----------------|------------------------|-------------------------|
| Previous Certificate | <i>Dated</i> | <i>Certificate No.</i> | <i>Laboratory</i> |
| | 25 January 2017 | TCRT17/1021 | ANV Measurement Systems |

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1165

Page 2 of 3 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|--------------------------------|-----------------------|--|
| SLM instruction manual title | NL-22 NL-32 Instruction Manual | | |
| SLM instruction manual ref / issue | 33625 09-06 | | |
| SLM instruction manual source | Manufacturer | | |
| Internet download date if applicable | N/A | | |
| Case corrections available | Yes | | |
| Uncertainties of case corrections | No | See comment on page 3 | |
| Source of case data | Manufacturer | | |
| Wind screen corrections available | Yes | | |
| Uncertainties of wind screen corrections | No | See comment on page 3 | |
| Source of wind screen data | Manufacturer | | |
| Mic pressure to free field corrections | Yes | | |
| Uncertainties of Mic to F.F. corrections | No | See comment on page 3 | |
| Source of Mic to F.F. corrections | Manufacturer | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | | Yes | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Lab Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 07 February 2018 | | |
| Calibrator cert. number | UCRT18/1145 | | |
| Calibrator cal cert issued by Lab. | 0653 | | |
| Calibrator SPL @ STP | 94.03 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1001.88 | Hz | Calibration check frequency |
| Reference level range | 30 - 120 | dB | |

Accessories used or corrected for during calibration - None

Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 22.41 | 22.67 | ± 0.30 °C |
| Humidity | 39.8 | 41.6 | ± 3.00 %RH |
| Ambient Pressure | 101.42 | 101.35 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|--|------|----|--------------------------|------|----|
| Initial indicated level | 94.0 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|--|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

Microphone replaced with electrical input device - UR = Under Range indicated

| Weighting | A | C | Z |
|--|---------|------|------|
| | 11.3 | 17.9 | 23.7 |
| | dB | dB | dB |
| | UR | UR | |
| Uncertainty of the electrical self generated noise ± | 0.12 dB | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by the International Organisation for Standards (ISO).

Comments

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1165

Page 3 of 3 Pages

If any of the "Uncertainties of" are set to NO above, then the following applies.

No information on the uncertainty of measurement, required by 11.7 of IEC 61672-3:2006, of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, or the manufacturer of the multi-frequency sound calibrator, or the manufacturer of the electrostatic actuator was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of the measurement of the adjustment data has therefore been assumed to be numerically zero for the purpose of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of IEC 61672-1:2002.

Calibrated by: A Patel

R 1

..... END

Additional Comments

None



CERTIFICATE OF CALIBRATION

Date of Issue: 22 May 2018

Certificate Number: TCRT18/1449

Issued by:

ANV Measurement Systems

Beaufort Court

17 Roebuck Way

Milton Keynes MK5 8HL

Telephone 01908 642846 Fax 01908 642814

E-Mail: info@noise-and-vibration.co.uk

Web: www.noise-and-vibration.co.uk

Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems

Page 1 of 3 Pages

Approved Signatory

K. Mistry

Customer TNEI Services Ltd
7th Floor
West One, Forth Banks
Newcastle Upon Tyne
NE1 3PA

Order No. 5001
Description Sound Level Meter / Pre-amp / Microphone / Associated Calibrator
Identification

| Manufacturer | Instrument | Type | Serial No. / Version |
|--------------|---------------------------------------|--------|----------------------|
| Rion | Sound Level Meter | NL-31 | 01283554 |
| Rion | Firmware | | 1.400 |
| Rion | Pre Amplifier | NH-21 | 29311 |
| Rion | Microphone | UC-53A | 315581 |
| Rion | Calibrator | NC-74 | 34536109 |
| | Calibrator adaptor type if applicable | | NC-74-002 |

Performance Class 1
Test Procedure TP 2.SLM 61672-3 TPS-49
Procedures from IEC 61672-3:2006 were used to perform the periodic test.
Type Approved to IEC 61672-1:2002 No Approval Number
If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2003
Date Received 21 May 2018 ANV Job No. TRAC18/05251
Date Calibrated 22 May 2018

The sound level meter submitted for testing has successfully completed the class 1 periodic tests of IEC 61672-3:2006, for the environmental conditions under which the tests were performed. However, no general statement or conclusion can be made about conformance of the sound level meter to the full requirements of IEC 61672-1:2002 because evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to demonstrate that the model of sound level meter fully conformed to the requirements in IEC 61672-1:2002 and because the periodic tests of IEC 61672-3:2006 cover only a limited subset of the specifications in IEC 61672-1:2002.

| Previous Certificate | Dated | Certificate No. | Laboratory |
|----------------------|-------------|-----------------|-------------------------|
| | 31 May 2017 | TCRT17/1315 | ANV Measurement Systems |

This certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1449

Page 2 of 3 Pages

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|--------------------------------|-----------------------|--|
| SLM instruction manual title | NL-21 NL-31 Instruction Manual | | |
| SLM instruction manual ref / issue | 32006 09-04 | | |
| SLM instruction manual source | Manufacturer | | |
| Internet download date if applicable | N/A | | |
| Case corrections available | Yes | | |
| Uncertainties of case corrections | No | See comment on page 3 | |
| Source of case data | Manufacturer | | |
| Wind screen corrections available | Yes | | |
| Uncertainties of wind screen corrections | No | See comment on page 3 | |
| Source of wind screen data | Manufacturer | | |
| Mic pressure to free field corrections | Yes | | |
| Uncertainties of Mic to F.F. corrections | No | See comment on page 3 | |
| Source of Mic to F.F. corrections | Manufacturer | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2002 | Yes | | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Lab Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 09 May 2018 | | |
| Calibrator cert. number | UCRT18/1502 | | |
| Calibrator cal cert issued by Lab. | ANV Measurement Systems | | |
| Calibrator SPL @ STP | 94.02 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1001.88 | Hz | Calibration check frequency |
| Reference level range | 30 - 120 | dB | |

Accessories used or corrected for during calibration - None
 Note - if a pre-amp extension cable is listed then it was used between the SLM and the pre-amp.

| Environmental conditions during tests | Start | End | |
|---------------------------------------|--------|--------|------------|
| Temperature | 24.00 | 23.95 | ± 0.30 °C |
| Humidity | 40.9 | 34.9 | ± 3.00 %RH |
| Ambient Pressure | 100.71 | 100.73 | ± 0.03 kPa |

Response to associated Calibrator at the environmental conditions above.

| | | | | | |
|--|------|----|--------------------------|------|----|
| Initial indicated level | 94.0 | dB | Adjusted indicated level | 94.0 | dB |
| The uncertainty of the associated calibrator supplied with the sound level meter ± | | | 0.10 dB | | |

Self Generated Noise This test is currently not performed by this Lab.

| | | | |
|--|-----|----|-------------|
| Microphone installed (if requested by customer) = Less Than | N/A | dB | A Weighting |
| Uncertainty of the microphone installed self generated noise ± | N/A | dB | |

Microphone replaced with electrical input device - UR = Under Range indicated

| Weighting | A | | | C | | | Z | | |
|--|------|----|--|------|----|--|------|----|--|
| | dB | UR | | dB | UR | | dB | UR | |
| | 11.4 | | | 17.1 | | | 22.7 | | |
| Uncertainty of the electrical self generated noise ± | | | | 0.12 | | | dB | | |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with the Guide to the Expression of Uncertainty in Measurement published by the International Organisation for Standards (ISO).

Comments

For the test of the frequency weightings as per paragraph 12. of IEC 61672-3:2006 the actual microphone free field response was used.

The acoustical frequency tests of a frequency weighting as per paragraph 11 of IEC 61672-3:2006 were carried out using an electrostatic actuator.

CERTIFICATE OF CALIBRATION



Certificate Number

TCRT18/1449

Page 3 of 3 Pages

If any of the "Uncertainties of" are set to NO above, then the following applies.

No information on the uncertainty of measurement, required by 11.7 of IEC 61672-3:2006, of the adjustment data given in the instruction manual or obtained from the manufacturer or supplier of the sound level meter, or the manufacturer of the microphone, or the manufacturer of the multi-frequency sound calibrator, or the manufacturer of the electrostatic actuator was published in the instruction manual or made available by the manufacturer or supplier. The uncertainty of the measurement of the adjustment data has therefore been assumed to be numerically zero for the purpose of this periodic test. If these uncertainties are not actually zero, there is a possibility that the frequency response of the sound level meter may not conform to the requirements of IEC 61672-1:2002.

Calibrated by: A Patel

R 1

..... END

Additional Comments

None



CERTIFICATE OF CONFORMANCE

Date of Issue 08 March 2018
Customer CONF031805
Certificate Number TNEI Services Ltd

| | Manufacturer | Type | Serial |
|--------------------------|---------------------|-------------|---------------|
| Sound Level Meter | Rion | NL-52 | 01176428 |
| Preamplifier | Rion | NH-25 | 76447 |
| Microphone | Rion | UC-59 | 12471 |

This is to certify that the instrument was tested and calibrated at the Manufacturer's factory according to their specification and that the product satisfied all the relevant requirements of the following Standards:

IEC 61672-1:2013 Class 1.

The instrument also received a functional check by ANV Measurement Systems prior to despatch in the UK, in accordance with our standard procedures.

Signed *Amrat C Patel* Position. Calibration Technician Date. 08 March 2018
A Patel

BEAUFORT COURT, 17 ROEBUCK WAY, MILTON KEYNES, MK5 8HL

☎ 01908 642846 📠 01908 642814

✉ info@noise-and-vibration.co.uk 🌐 www.noise-and-vibration.co.uk



CERTIFICATE OF CONFORMANCE

Date of Issue 20 June 2018
Customer TNEI Services Ltd
Certificate Number CONF061809

| | Manufacturer | Type | Serial |
|--------------------------|---------------------|-------------|---------------|
| Sound Level Meter | Rion | NL-52 | 000386739 |
| Preamplifier | Rion | NH-25 | 76889 |
| Microphone | Rion | UC-59 | 12362 |

This is to certify that the instrument was tested and calibrated at the Manufacturer's factory according to their specification and that the product satisfied all the relevant requirements of the following Standards:

IEC 61672-1:2013 Class 1.

The instrument also received a functional check by ANV Measurement Systems prior to despatch in the UK, in accordance with our standard procedures.

Signed *A Patel* Position. Calibration Technician Date. 20 June 2018
A Patel

BEAUFORT COURT, 17 ROEBUCK WAY, MILTON KEYNES, MK5 8HL

☎ 01908 642846 📠 01908 642814

✉ info@noise-and-vibration.co.uk 🌐 www.noise-and-vibration.co.uk



CERTIFICATE OF CONFORMANCE

Date of Issue 20 June 2018
Customer TNEI Services Ltd
Certificate Number CONF061806

| | Manufacturer | Type | Serial |
|--------------------------|---------------------|-------------|---------------|
| Sound Level Meter | Rion | NL-52 | 00386758 |
| Preamplifier | Rion | NH-25 | 76908 |
| Microphone | Rion | UC-59 | 12755 |

This is to certify that the instrument was tested and calibrated at the Manufacturer's factory according to their specification and that the product satisfied all the relevant requirements of the following Standards:

IEC 61672-1:2013 Class 1.

The instrument also received a functional check by ANV Measurement Systems prior to despatch in the UK, in accordance with our standard procedures.

Signed Amrat C. Patel Position. Calibration Technician Date. 20 June 2018
A Patel

BEAUFORT COURT, 17 ROEBUCK WAY, MILTON KEYNES, MK5 8HL

☎ 01908 642846 📠 01908 642814

✉ info@noise-and-vibration.co.uk 🌐 www.noise-and-vibration.co.uk



CERTIFICATE OF CONFORMANCE

Date of Issue **20 June 2018**
Customer **TNEI Services Ltd**
Certificate Number **CONF061807**

| | Manufacturer | Type | Serial |
|--------------------------|---------------------|-------------|---------------|
| Sound Level Meter | Rion | NL-52 | 000386759 |
| Preamplifier | Rion | NH-25 | 76909 |
| Microphone | Rion | UC-59 | 12756 |

This is to certify that the instrument was tested and calibrated at the Manufacturer's factory according to their specification and that the product satisfied all the relevant requirements of the following Standards:

IEC 61672-1:2013 Class 1.

The instrument also received a functional check by ANV Measurement Systems prior to despatch in the UK, in accordance with our standard procedures.

Signed *Arvind C. Patel* Position. Calibration Technician Date. 20 June 2018
A Patel

BEAUFORT COURT, 17 ROEBUCK WAY, MILTON KEYNES, MK5 8HL

☎ 01908 642846 📠 01908 642814

✉ info@noise-and-vibration.co.uk 🌐 www.noise-and-vibration.co.uk



CERTIFICATE OF CONFORMANCE

Date of Issue 20 June 2018
Customer TNEI Services Ltd
Certificate Number CONF061811

| | Manufacturer | Type | Serial |
|--------------------------|---------------------|-------------|---------------|
| Sound Level Meter | Rion | NL-52 | 00386760 |
| Preamplifier | Rion | NH-25 | 76910 |
| Microphone | Rion | UC-59 | 12778 |

This is to certify that the instrument was tested and calibrated at the Manufacturer's factory according to their specification and that the product satisfied all the relevant requirements of the following Standards:

IEC 61672-1:2013 Class 1.

The instrument also received a functional check by ANV Measurement Systems prior to despatch in the UK, in accordance with our standard procedures.

Signed *Anvat C. Patel* Position. Calibration Technician Date. 20 June 2018
A Patel

BEAUFORT COURT, 17 ROEBUCK WAY, MILTON KEYNES, MK5 8HL

☎ 01908 642846 📠 01908 642814

✉ info@noise-and-vibration.co.uk 🌐 www.noise-and-vibration.co.uk



CERTIFICATE OF CONFORMANCE

Date of Issue 20 June 2018
Customer TNEI Services Ltd
Certificate Number CONF061808

| | Manufacturer | Type | Serial |
|--------------------------|---------------------|-------------|---------------|
| Sound Level Meter | Rion | NL-52 | 000386761 |
| Preamplifier | Rion | NH-25 | 76911 |
| Microphone | Rion | UC-59 | 12788 |

This is to certify that the instrument was tested and calibrated at the Manufacturer's factory according to their specification and that the product satisfied all the relevant requirements of the following Standards:

IEC 61672-1:2013 Class 1.

The instrument also received a functional check by ANV Measurement Systems prior to despatch in the UK, in accordance with our standard procedures.

Signed Amrat C Patel Position. Calibration Technician Date. 20 June 2018
A Patel

BEAUFORT COURT, 17 ROEBUCK WAY, MILTON KEYNES, MK5 8HL

☎ 01908 642846 📠 01908 642814

✉ info@noise-and-vibration.co.uk 🌐 www.noise-and-vibration.co.uk